



EC TYPE EXAMINATION CERTIFICATE (G-MODULE CERTIFICATE)

Number: N-10/2008 dated 24.11.2008

Notified body number 0431

Issued by:	Norwegian Metrology Service (Justervesenet) Fetveien 99, N-2007 Kjeller NORWAY
Issued to:	ECS Automasjon AS Postboks 508, 4349 Bryne NORWAY
In respect of:	One discontinuous totalizing automatic weighing instrument
In accordance with:	Act on Metrology and Verification dated 2007.12.20 no.1723 and corresponding regulation no. 1737 dated 2007.12.21 in connection with Directive 2004/22/EC of the European Parliament and Council of March 31, 2004 on measuring instruments (MID)
Applicable essential requirements: The instrument is placed in:	Annex 1 and Annex MI-006 Welcon Protein AS Varbergveien 61, 4370 Egersund NORWAY
Type designation:	Sartorius X5-BULK
Characteristics:	
Accuracy class:	0,2
Scale interval, d:	1 kg
Minimum totalized load, Σ_{min}	1000 kg
Maximum capacity:	2000 kg
Temperature range:	-10 °C /40 °C
Mechanical class:	M1
Electromagnetic environment class:	E2
Valid until:	24.11.2018
Number of pages:	1 including annex 1 consisting of 7 pages
Reference No.:	07/446-T-4306 (07/446 is the project number and 4306 is the application number).
Date of issue:	24.11.2008
Approved by :	Processed By
Mr. Knut Lindløv Director, head of Legal Metrology Norwegian Metrology Service	André Evenstad Senior engineer Norwegian Metrology Service