

EC TYPE EXAMINATION CERTIFICATE (G-MODULE)
Number: N-05/962 dated 2009-02-13, Revision 0
Notified body number 0431

Issued by: Norwegian Metrology Service (Justervesenet)
Fetveien 99, N-2007 Kjeller
NORWAY

Issued to: Mittett Tekniske AS
Post BOX 124, 7566 VIKHAMMER
Norway

In respect of: One electronic automatic gravimetric filling
Instrument

Instrument is placed: Steinkjer Kornsilø AS, 7700 Steinkjer, Norway

In accordance with: Act on regulation no. 1723 dated 2007.12.20 on
measuring units and measurement and corresponding
regulation no. 1753 dated 2007.12.21 in connection
with Directive 2004/22/EC of the European Parliament
and Council of March 31, 2004 on measuring
instruments (MID)

Applicable essential requirements: Annex 1 & Annex MI-6

Type designation: IT6000

Characteristics:


Accuracy class: X (1)
Maximum capacity: 40 kg
Temperature range: -10 °C /40°C
Mechanical class: M2
Electromagnetic environment class: E2


Valid until: 2019-02-13

Number of pages: 1 including annex 1 consisting of 7 pages

Reference No.: 05/962-T4213 (05/926 is the project number and 4213
is the *application number*).

Date of issue: 2009-02-13

Approved by :

Knut Lindløv
Director, head of Legal Metrology
Norwegian Metrology Service

Processed By

Hossein Piltan
Senior engineer