



EC TYPE EXAMINATION CERTIFICATE (G-MODULE CERTIFICATE)

Number: N-09/2008 dated 21.11.2008

Notified body number 0431

Issued by:	Norwegian Metrology Service (Justervesenet) Fetveien 99, N-2007 Kjeller NORWAY
Issued to:	Mittet Tekniske AS Postboks 124, 7566 Vikhammer NORWAY
In respect of:	One automatic gravimetric filling instrument
In accordance with:	Act on Metrology and Verification dated 2007.12.20 no.1723 and corresponding regulation no. 1747 dated 2007.12.21 in connection with Directive 2004/22/EC of the European Parliament and Council of March 31, 2004 on measuring instruments (MID)
Applicable essential requirements:	Annex 1 and Annex MI-006
The instrument is placed in:	Holla Metall avd Fesil ASA 7200 Kyrksæterøra NORWAY
Type designation:	IND560
Characteristics:	
Accuracy class:	X(1)
Scale interval, e:	1 kg
Minimum capacity:	200 kg
Maximum capacity:	1500 kg
Temperature range:	-10 °C /40 °C
Mechanical class:	M1
Electromagnetic environment class:	E2
Valid until:	21.11.2018
Number of pages:	1 including annex 1 consisting of 6 pages
Reference No.:	06/387-T-4240 (06/387 is the project number and 4240 is the application number).
Date of issue:	21.11.2008
Approved by :	Processed By
 Mr. Knut Lindløv	 André Evenstad
Director, head of Legal Metrology Norwegian Metrology Service	Senior engineer Norwegian Metrology Service