

Justervesenet

Norwegian Metrology and Accreditation Service

Certificate of EC type-approval no N-02/98-0431 of a non-automatic weighing instrument

Issued by Norwegian Metrology and Accreditation Service
Notified Body Number 0431

In accordance with the requirements of the Non-automatic Weighing Instrument
Regulation 10/94 which implements the Council Directive
90/384/EEC in Norway

Issued to Pols Electronics Ltd
Sindragata 10
400 Isafjorder
Iceland

In respect of a non-automatic single or multi-interval weighing instrument
designated V-182 or V-185, which are a modular system of
weighing components based on the indicator P-15 and variants
of modules of load receptors, load cells and peripheral
equipment.

Accuracy class	III
Maximum capacity:	Max \geq 1500 g
Minimum capacity:	Min = 20 * e
Verification scale interval:	e \geq 0.2 g
Maximum number of scale intervals:	n = 3000

The necessary data (principal characteristics, securing, functioning etc.) for identification purposes and approval conditions are set out in the descriptive annex to this certificate. The annex comprises 8 pages.

Knut Lindløv
Director Legal Metrology

Bente Lie

Date: 15 October 1998
Valid until: 15 October 2008

Avdelinger / Departments :	Adresse / Address :	Tel. : (+47) 64 84 84 84	Bankkonto : 7029 05 03858
Nasjonale normaler	Fetveien 99	Fax : (+47) 64 84 84 85	Postgirokonto : 0809 4050902
Legal metrologi	2007 KJELLER	E-mail : postmottak@	Org.nr. : 874 761 192
Norsk Akkreditering	NORWAY	justervesenet.dep.telemex.no	