

**EC TYPE EXAMINATION CERTIFICATE**  
**Number: N-08/536 dated 2009-05-07, revision 0**  
**Notified body number 0431**

Issued by: Norwegian Metrology Service (Justervesenet)  
Fetveien 99, N-2007 Kjeller  
NORWAY

Issued to: Espera-Werke GmbH  
Moltkestrasse 17-33  
D-47058 Duisburg, Germany

In respect of: Automatic catchweighing instrument

In accordance with: Act on regulation no. 1723 dated 2007.12.20 on  
measuring units and measurement and corresponding  
regulation no. 1742 dated 2007.12.21 in connection  
with Directive 2004/22/EC of the European Parliament  
and Council of March 31, 2004 on measuring  
instruments (MID)

Applicable essential requirements: Annex 1 & Annex MI-006


Type designation: ES 7xyz


**Characteristics:**

Accuracy class: XIII(1) and Y(a)  
Minimum capacity: 100g for XIII(1) and 20e for Y(a)  
Number of verification scale interval(n): 3000  
Preset tare, subtractive: < Max  
Temperature range: -10 °C /40 °C  
Belt speed: < 70m/m  
Mechanical class: M1  
Electromagnetic environment class: E1

Valid until: 2019-05-07  
Number of pages: 1 including annex 1 consisting of 6 pages  
Reference No.: 08/539  
Date of issue: 2009-05-07

Approved by : Processed By

  
Knut Lindløv  
Director, head of Legal Metrology,  
Norwegian Metrology Service

  
Hossein Piltan  
Senior engineer