

EC UNIT VERIFICATION CERTIFICATE (G-module)
Number: N-07/334 dated 2009-11-25, revision 0
Notified body number 0431

Issued by: Norwegian Metrology Service (Justervesenet)
Fetveien 99, N-2007 Kjeller
NORWAY

Issued to: Xstrata Nikkelverk AS, Vesterveien 31
4613 Kristiansand, Norway

In respect of: 6 automatic catch weighing instruments with serial no.
Pos. 5705E, Pos.4204, Pos. 4206, pos. 4122, Pos.4312
and Pos.4327F

In accordance with: Act on regulation no. 1723 dated 2007.12.20 on
measuring units and measurement and corresponding
regulation no. 1742 dated 2007.12.21 in connection
with Directive 2004/22/EC of the European Parliament
and Council of March 31, 2004 on measuring
instruments (MID)

Applicable essential requirements: Annex 1 & Annex MI-006

Type designation: Weighing module: SIWAREX FTA
Indicating device : PLC of type TP-panel

Characteristics:

Accuracy class: Y(a)
Minimum capacity: 20 kg
Maximum capacity: 3000 kg
Verification scale interval, e: 1 kg
Number of verification scale interval(n): 3000
Preset tare, subtractive: ≤Max
Temperature range: -10 °C /40 °C
Electromagnetic environment class: E2

Valid until: 2019-11-25
Number of pages: 1 including annex 1 consisting of 7 pages
Reference No.: 07/334 which is project number
Date of issue: 2009-11-25

Approved by : 
Hans Arne Frøystein
Director, head of National laboratory

Processed By 
Hossein Piltan
Type approval senior engineer