

EC TYPE EXAMINATION CERTIFICATE
Number: N-07/359 dated 2010-04-23 revision 0
Notified body number 0431

Issued by: Norwegian Metrology Service (Justervesenet)
Fetveien 99, N-2007 Kjeller
NORWAY

Issued to: PLS-teknikk a/s
Tvetenveien 162, 0671 Oslo
NORWAY

In respect of: Discontinuous totalizing automatic weighing instruments

In accordance with: Act on Metrology and Verification dated 2007.12.20 no.1723 and corresponding regulation no. 1737 dated 2007.12.21 in connection with Directive 2004/22/EC of the European Parliament and Council of March 31, 2004 on measuring instruments (MID)

Applicable essential requirements: Annex 1 and Annex MI-006
The instruments are placed in:

Type designation: Panther

Characteristics:

Accuracy class: 0.2

Scale interval (d): $d \geq 0,2 \text{ kg}$

Number of scale internal, n: $n \leq 3000$

Minimum totalized load, (Σ_{\min}): $\Sigma_{\min} \geq 200 \text{ kg}$

Maximum capacity: $\leq 5000 \text{ kg}$

Temperature range: $-10 \text{ }^\circ\text{C} / 40 \text{ }^\circ\text{C}$

Electromagnetic environment class: E2


Valid until: 2020-04-15

Number of pages: 1 including annex 1 consisting of 7 pages


Reference No.: **07/359-T-4301** (07/359 is the project number and 4301 is the application number).

Date of issue: 2010-04-23

Approved by :


Mr. Hans Arne Frøystein
Director, head of National Laboratory

Processed By


Hossein Piltan
Senior engineer